




Agilent 86100A/B/C Infiniium DCA Series

Firmware Release Notes



Firmware Release Notes

© Copyright Agilent Technologies
Company 2010
All Rights Reserved. Reproduction,
adaptation, or translation without
prior written permission is prohib-
ited, except as allowed under copy-
right laws.

July 2010
Second Edition
Made in USA.

Agilent Technologies Company
Digital Test Division
1400 Fountaingrove Parkway
Santa Rosa, CA 95403, USA

Notice.

The information contained in this document is subject to change without notice. Companies, names, and data used in examples herein are fictitious unless otherwise noted. Agilent Technologies makes no warranty of any kind with regard to this material, including but not limited to, the implied warranties of merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Restricted Rights Legend.

Use, duplication, or disclosure by the U.S. Government is subject to restrictions as set forth in subparagraph (c) (1) (ii) of the Rights in Technical Data and Computer Software clause at DFARS 252.227-7013 for DOD agencies, and subparagraphs (c) (1) and (c) (2) of the Commercial Computer Software Restricted Rights clause at FAR 52.227-19 for other agencies.

Contents

86100C-Only Firmware Release Notes 4

A.09.01 Release Notes, July, 2010 (Differences from Release A.09.00)	4
A.09.00 Release Notes, June 10, 2010 (Differences from Release A.08.10)	4
A.08.10 Release Notes, April 8, 2009 (Differences from Release A.08.00)	4
A.08.00 Release Notes, June 18, 2008 (Differences from Release A.07.00)	5
A.07.00 Release Notes, March 15, 2007 (Differences from Release A.06.01)	8
A.06.01 Release Notes, September 21, 2006 (Differences from Release A.06.00)	9

86100C/86100B/86100A Firmware Release Notes 11

A.05.00 Release Notes, February 16, 2005 (Differences from Release A.04.20)	11
A.04.20 Release Notes, June 9, 2005 (Differences from Release A.04.11)	11
A.04.11 Release Notes, December 10, 2004 (Differences from Release A.04.10)	13
A.04.10 Release Notes, November 1, 2004 (Differences from Release A.04.01)	14
A.04.01 Release Notes, June 23, 2004 (Differences from Release A.04.00)	16
A.04.00 Release Notes, February 19, 2004 (Differences from Release A.03.05)	16
A.03.04 Release Notes, October 10, 2002 (Differences from Release A.03.03)	19
A.03.03 Release Notes, July 18, 2002 (Differences from Release A.03.03)	19

86100A-Only Firmware Release Notes 21

A.03.01 Release Notes, February 15, 2002 (Differences from Release A.02.20)	21
A.02.20 Release Notes, August 2, 2001 (Differences from Release A.02.10)	22
A.02.10 Release Notes, May 1, 2001 (Differences from Release A.02.01)	23
A.02.01 Release Notes, January 30, 2001 (Differences from Release A.02.00)	24
A.02.00 Release Notes, November 29, 2000 (Differences from Release A.01.22)	24
A.01.22 Release Notes, September 6, 2000 (Differences from Release A.01.21)	26
A.01.21 Release Notes, August 9, 2000 (Differences from Release A.01.20)	26
A.01.20 Release Notes, July 11, 2000 (Differences from Release A.01.13)	26
A.01.13 Release Notes, May 2, 2000 (Differences from Release A.01.12)	27

86100C-Only Firmware Release Notes

A.09.01 Release Notes, July, 2010 (Differences from Release A.09.00)

General Defects Fixed

- Improved the behavior of Autoscale and Jitter mode when using the precision timebase with spread-spectrum clocks.

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.09.00 Release Notes, June 10, 2010 (Differences from Release A.08.10)

New Feature List

- Support for new 86115D 20 GHz optical module.
- Seven new standard masks. Five cover 802.3, Revision D3.1 and two cover FC-PI5, Revision 1.0.
- Eye tuning in Eye/Mask mode adds the ability to tune a device while simultaneously watching changes to the eye diagram and eye diagram measurements.
- Improved online help system with search capability.

General Defects Fixed

- Jitter Database file (.CSV) now correctly saves the TI CDF one level data.
- Several eye-measurements bugs were fixed.
- Increased number of digits in touchstone format.
- Fixed impedance values when saving differential and common mode s-parameters in touchstone files.
- Fixed a crash that occurred after vertical calibration if histograms are turned on.
- Increased maximum average count in TDR mode.

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.08.10 Release Notes, April 8, 2009 (Differences from Release A.08.00)

New Feature List

- Clock Recovery Up To 14.2 Gb/s: The 86108A Precision Waveform Analyzer and 83496B Clock Recovery modules now run from 50 Mb/s to 14.2 Gb/s, allowing customers to recover

- embedded clocks from optical and electrical signals up to 16X Fibre Channel.
- Support for 86105D Optical/Electrical Module: The 86105D is a new module supporting 8X and 16X Fibre Channel as well as 10 Gb Ethernet and 10 G SONET/SDH line rates.
- Improved Clock Recovery Support For SATA: The upper limit for the loop bandwidth (LBW) has been increased to line rate/250, allowing 6 MHz LBW required for SATA I signals (1.5 Gb/s). The absolute LBW limit remains at 10 MHz.
- J2 and J9 TJ Measurements: J2 and J9 are new Total Jitter (TJ) measurements from standards requiring stressed eye analysis for optical and electrical receiver compliance.
- TJ Upper Limit Increased To 10E-1: The upper limit for the BER used in of Total Jitter (TJ) measurement BER has been increased from 1E-3 to 1E-1.
- I-Q Diagram in Color Grade: The math function ?gVersus?h allows color-graded I-Q diagrams, a key insight tool to visualize the I/Q signals after demodulation of a complex optical modulation.
- Remote Skew Control For 86118A-H01: New GP-IB commands now allow remote skew control for option H01 of the 86118A, a dual remote sampling head with over 70 GHz electrical bandwidth.
- Additional Trigger Divide Ratio: 1:55 has been added to support more applications using subrate clocks to trigger the oscilloscope.
- Additional Loop Bandwidth Value: Added 637.5 kHz to "Select from List" for target loop bandwidth (LBW) of the clock recovery.
- Calibration Timeout Increased to 30 Seconds: During most calibrations the hardware reacts within a fraction of a second. External programs, such as anti-virus and anti-spyware software, may keep the CPU busy for seconds. The long timeout ensures successful calibration in most if not all cases.

General Defects Fixed

- Fixed corner case where too much jitter was reported
- Fixed issue: jitter algorithm related to transition density that resulted in over-reported DJ/TJ numbers.
- Fixed occasional ghost error message (?gReset PTB?h) when no precision timebase was installed.

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.08.00 Release Notes, June 18, 2008 (Differences from Release A.07.00)

New Feature List

- Support for the new 86108A Precision Waveform Analyzer module. The 86108A is a double wide module that features two electrical inputs, clock recovery and a precision timebase. Previously, the combination of these required three separate modules. Now they are incorporated into a single, double wide module.
- Support for reference receiver options 86116C-025 and 86116C-040. The 86116C has been replaced with the 86116C-040 and 86116C-025. Option 040 is a direct replacement for the legacy 86116C as a reference receiver for 39.8 GB/s and 43.0 Gb/s NRZ signals. Option 025 offers reference receiver capability for 17, 25.8 and 27.7 Gb/s NRZ signals which were previously not available in the 86116C.
- Support for the 86100CU-400 PLL / Jitter Transfer Analysis option.
- Support for the Bulk Optics change on the 83496B Clock Recovery Modules.
- 86108A Setup Wizard to help configure the 86108A module. This wizard has preset

- configurations for a number of use cases to make the configuration as simple as possible.
- Advanced Clock Data Recovery Setup for the 86108A allowing user to specify Type 2 Transition Frequency (peaking) of the loop.
 - Support for MATLAB measurements with scripts created by the user. Now the user can make custom measurements on the DCA. These measurements run just like internal measurements, with results on the Measurement Tab, including statistics.
 - Improved jitter measurement accuracy for very large jitter. Increased limit for maximum amount of measureable jitter.
 - Added Large Jitter Mode for extending jitter measurement capabilities. This mode is enabled when the amount of jitter exceeds the transition time of the signal. As with the new edge model, Large Jitter Mode will also increase the accuracy of large jitter measurements and increases the maximum amount of measureable jitter.
 - New S-parameter plots, phase and group delay, for 86100C-202 Enhanced Impedance and S-Parameter Software package.
 - Auto Mask Margin is a new enhancement to the Eye Mode Mask Margin feature which will search for a scaled mask that matches a user specified bit error rate.
 - Auto Relock is a new feature for clock recovery modules. Previously, when a recovered clock had lost lock, the user would have to manually relock. The 8.0 release now has the functionality to do this automatically.
 - Users can now send remote commands to the instrument directly over Ethernet LAN. GPIB is no longer required for remote operation.
 - Increased the maximum record length from 4,096 to 16,384.
 - Now ?gPattern Lock?h on 86100Cs with option 001 removes the 4 ns discontinuities on long spans (no time tears) as long as the trigger delay is at or close to its minimum value. In this case the instrument synchronizes the timebase to the pattern and force samples to be taken at the correct locations.
 - Increased the maximum timebase range while using Pattern Lock from 250.0 ns/div to 250.0 us/div.
 - Increased the speed of the Pattern Waveform Save feature.
 - Reduced the minimum points/bit of Pattern Waveform Save from 16 to 1.
 - Added support for Trigger Divide Ratios of 1:18 and 1:36.

General Defects Fixed

- XY Verbose values now correspond to the values being shown on the display with account for the reference plane.
- Fixed crash that would happen when attempting to run the Linear Forward Equalizer from a MATLAB script which was not available.
- Fixed issues with changing the trigger source while precision timebase is on.
- Fixed two issues referring to DTA messages complaining about precision timebase issues. One would complain about the time reference being out of band when the precision timebase was not on, and the other message would ask user to restore reference clock after long continuous usage of Pattern Lock and precision timebase.
- Fixed a issue where bad data would get reported on clipped signals with precision timebase.
- Fixed an issue where math signals were not auto scaling properly.
- If TDR calibration is valid and then a vertical cal later becomes invalid/recommended, now a message will tell user TDR calibration is recommended.
- Fixed bug where :DISK:MDIRectory command would give error message upon proper use of command.
- The minimum, maximum, and average impedance measurements can now be applied to math functions as well as channels and responses.
- Fixed issues with using Pattern Lock and precision timebase with a sub-rate reference clock.

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.07.00 Release Notes, March 15, 2007 (Differences from Release A.06.01)

New Features List

- The following features are available only with Option 300, Advanced Amplitude Analysis/RIN/Q-factor (which requires Option 200, Enhanced Jitter Analysis Software and 86100C Option 001, Enhanced Trigger): Amplitude Analysis, Separation and display of amplitude interference components (similar to jitter mode), RIN/RINoma, Q factor
- Added Optical Modulation Amplitude measurement to oscilloscope mode.
- Support for 83496B clock recovery module with the capability to lock on data streams that employ a Spread Spectrum Clock (SSC). Enables phase noise analysis on clocks.
- Support for Option 83496A-UAB, upgrade to 83496B capability.
- Support for new 86116C high speed optical and electrical module. The 86116C offers reference receiver performance for the OC768 standard.
- Upper limit on LBW increased to 24M/n , which allows wider bandwidth for signals with .5 transition density.
- Added queries DISK:TFILE? and DISK:BFILE? To transfer text and binary files over GPIB
- New setting can automatically turn on Quick Measure when entering Eye/Mask mode or Oscilloscope mode.

General Defects Fixed

- The queries time:pos?, time:range?, time:scale?, wav:xdis?, wav:xran? Now properly return zoomed parameters.
- Waveform display for TDR common mode response now update properly after zoom.
- ER calibration delta temp now uses the module temperature.
- RZ One and Zero level measurements now make measurement at the peak of the pulse instead of the center.
- DISPlay:DATA? And DISK:SIMage with GRATicule parameter will not include the mask test results tab in the screen image.
- Saving pattern waveform now uses *OPC when finished.
- Fixed Differential TDR calibration update problem when changing timebase settings.
- Repaired an issue that could occur when trying to load a Calibration file, that does not exist, over GPIB.
- Load jitter data base is no longer available outside jitter mode.
- Horizontal Auto Skew functionality improved.
- Zoom in DDJ vs. Bit plot is now properly limited to 2 bits.
- Repaired an issue that could occur when using LFE or Matlab with no modules present.
- Mask test always includes hits from final acquisition when stop is pressed during mask test.

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.06.01 Release Notes, September 21, 2006 (Differences from Release A.06.00)

New Feature List

- Added eye masks for 10G BASE LRM, 1000Base-LX10, 100Base-BX10, ITU g.984.2

General Defects Fixed

- Corrected the 100Base-BX/LX mask scale to now have Delta T set to 8E-9 seconds
- Function<N>:Vertical? Now returns correct value.
- Differential TDR deskew is correctly locked out when calibration is active.
- Missing specification for max input for 83496A included in help file.
- Missing note on specification for 83491/2/3 included in help file.
- Changed module vertical calibration messages for 86118A modules to say "Connect all necessary cables and adaptors for measurement to the <Left | Right> module.
- Addressed MEAS:TDR:MAX, MIN, AVE queries after default setup causing a crash.
- No longer issues a GPIB hardware missing error using autoscale after tdr/tdt cal.
- LTER? Now correctly sets bit 0 (COMP) when acquisition is complete.
- :TDR:STIMulus:EXTernal:POLarity? now returns correct data when longform is on.
- Fixed a problem when reading in a pattern waveform csv file written by hand.
- Fixed intermittent application crash during certain TDR mode recall setups.
- Fixed crash that happened with :MEAS:VTIM? .000001 in TDR mode right after default setup.
- Fixed a bug causing probe calibrations to fail (Traces may be displayed during calibrations).
- Added special options H93 and H95 for 5.00Gb/s and 6.25Gb/s to the 86105C.

Software Option Packages (#200 and #202) Defects Fixed

- Fixed problems with jitter plots after default setup (#200).
- Fixed problem where PJ is greater than DJ in some cases. (Still possible for PJ to be greater than DJ if DJ/PJ small enough since they are calculated from different histograms) (#200).
- Added new S-parameter specs to TDR (#202).
- Fixed defect preventing the save of differential or common mode S-Parameters from remote (#202).

Known Issues

- PLTS customers should not upgrade to A.06.00 and above software at this time since the executable properly. PLTS customers should use DCA firmware version A.04.11.
- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.
- 86100A/B Mainframes are not supported on DCA Firmware Revision A.06.00 and above.

A.06.00 Release Notes, December 15, 2005 (Differences from Release A.04.20)

New Feature List

- Added Enhanced Impedance and S-parameters (#202)
- Support for S-parameter results for single-ended (S11, S21, S12, and S22), differential (SDD11, SDD21, SDD22, SDD12), and common modes (SCC11, SCC21, SCC22, SCC12).
- S-parameter display graphs feature markers, graph formatting, Autoscale, and print support
- S-parameter is calculated between two time periods which allows customers to using time gating before converting to S-parameter
- Corrected impedance profile for single-ended and differential signals
- Allows export of S-parameter measurements into either .s1p or .s2p Touchstone data format for import into circuit simulators
- Support for 86100C Mainframe
- New TDR User Interface for measurement setup and display
- Improved TDR Calibration that supports time base changes, time delays, and zooming into data
- Guided TDR Calibration Procedure to improve usability
- Support of external TDR edge accelerators in TDR Mode
- Extensive support for math functions like derivatives, constant multiplications, add, subtract, and scaling
- Maximum/Minimum/Average Measurements in TDR Mode
- Export of time data in X & Y format
- Display of ER Cal factor
- Changed from hardware graphics to software graphics

General Firmware Defects Fixed in A.06.00

- Increased maximum allowable time base delay in TDR mode to 100 ns.
- 86100B + 54754A previously did not invert white trace color on print. The waveform color is set to gray during printing to prevent a white waveform trace on a white background.
- Repaired 86106B not successfully completing Autoscale with filter on
- Resolved ?g:DISK:SIM 'filename.jpg' ?gcausing instrument to hang when file specified is read-only. Added error checking before attempting to save file.
- Phase is correct now for reference clock frequencies between 25 GHz and 39 GHz on the 40 GHz input of an 86107A Option 40 module.
- Added a remote command to raise the jitter shade :DISPlay:JITTer:SHADe
- Allow TDR mode to accept maximum time base delay up to 100ns
- Extinction Ratio calibration factor is always loaded into dialog boxes
- After 86100C power cycle, TDR mode is recalled properly

Known Issues

- PLTS customers should not upgrade to A.06.00 software at this time since the executable will fail to run properly. PLTS customers should use DCA firmware version A.04.11.
- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.
- 86100A/B Mainframes are not supported on DCA Firmware Revision A.06.00

86100C/86100B/86100A Firmware Release Notes

A.05.00 Release Notes, February 16, 2005 (Differences from Release A.04.20)

New Feature List

- Support for 86100A/B/C Mainframe
- Maximum/Minimum/Average Measurements in TDR Mode
- Export of waveform data in X & Y format
- Display of ER Cal factor
- Added the following masks (100BASE-BX_LX10.msk, SATA 1.5Gb TX 250 cycles.msk, SATA 1.5GB TX 5 Cycles.msk)

General Firmware Defects Fixed in A.05.00

- 86100B + 54754A previously did not invert white trace color on print. The waveform color is set to gray during printing to prevent a white waveform trace on a white background.
- Repaired 86106B and other modules with un-labeled filter rates not successfully completing Autoscale when filter is enabled
- Phase is correct now for reference clock frequencies between 25 GHz and 39 GHz on the 40 GHz input of an 86107A Option 40 module.
- Repaired inconsistent pattern waveform loading behavior
- Extinction Ratio calibration factor is always loaded into dialog boxes
- Removed stray entry out of range message at end of ER calibration
- After 86100C power cycle, TDR mode is recalled properly
- Removed the ability to load and store TDR/TDT calibration data since not all instrument states were being saved properly.
- Allowed TDT normalization to work with external TDR stimulus
- Repaired a crash if remote commands (:MEAS:TDR: MAX MIN AVE) was sent without putting the oscilloscope in TDR mode first.

Known Issues

- PLTS customers should not upgrade to 5.0 software at this time since the executable will fail to run properly. PLTS customers should use DCA firmware version 4.11.
- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.04.20 Release Notes, June 9, 2005 (Differences from Release A.04.11)

New Feature List

- Support for 83496A Clock Recovery Module
- Supports Picosecond Pulse Lab 4020 Single ended TDR sources and Picosecond Pulse Lab 4022 Differential TDR/T Sources. Provides external common and differential mode

normalization to enable faster edges.

- Support for InfiniiMax II Series Probe
- Enables calibration debugging for module vertical calibration.
- Adds a method for stabilizing RJ value for jitter measurements. Remote commands added for RJ stabilization capability (#200).
- Determines linear feed forward equalizer tap values automatically (#201)
- Adds support for 8xFibre Channel filter rate (86105C #197)

General Firmware Defects Fixed in 4.20

- Repaired probe calibration date disappearing after autoscale or probe calibration on other channel.
- Fixed autoscale with no channels turned on does nothing. Autoscale now displays a dynamic text area message when autoscale is performed with no channels activated.
- Repaired autoscale with limit test activated causing software to hang.
- In TDR mode, fixed (:TIM:RANG 6.0E-001 and :MARK:STATe X1Y1,MANual,TRACk) remote commands sent quickly in succession causing a software crash.
- In TDR mode, resolved transducer settings are not always correct. Channel units picks are now fixed.
- In TDR mode, delta marker units track channel units, not marker units. This has been resolved.
- Repaired TDR normalization problem with center time reference. Responses are now being recalled even if reference is set to 'Center'.
- In TDR mode, resolved normalization staying valid after channel vertical calibration is performed.
- Provided excess reactance measurement which was not previously available for differential TDR responses.
- Added a remote query to see if TDR normalization is valid over GPIB.
- In TDR mode, fixed Ohms scale not available on Channel 2 after default setup
- Repaired extinction ratio calibration failing after precision time base is enabled. Extinction ratio calibration could fail with precision time base on if the horizontal scale was set for less than one period of the reference clock on screen. The calibration has been modified so that at least one period of the reference clock is on screen when dark level data is measured.
- Resolved oscilloscope hanging with 86107A when limit test is turned on. Disabled acquisition limit when setting the precision time base time reference. Hang no longer occurs.
- Resolved broken time base calibration. Hardware was left in an inconsistent setting relative to the control system, causing the next calibration to fail.
- Repaired ASSERTS when dialog is opened and DCA application is minimized. When minimized, application will be maximized on all front panel events (button press or knob movement) before event is processed.
- Repaired turning on of pattern lock after undo default setup with 86107A. Changed how the valid setup is built to account for priorities of controls for pattern lock and precision time base.
- Resolved application hang from Setup file load with divided trigger on. Reorganized acquisition limit display update to prevent excessive Windows messages from being generated and added code to remote commands to gracefully handle failure if occurs when placing windows messages. Hang no longer happens.
- Repaired release assertion during undo default setup. Fixed interaction between recall setup, autoscale, and the auto-record length control.
- Fixed combination of TIM:BRAT + :TRIG:RBIT + SYST:MODE JITT causing application

exit

- Resolved STOP CHAN1 :TIME:RANG:MEAS:VTIM? causes abrupt application exit
- Repaired MEASure:CGR:SOUR CGMemory causing application hang
- Resolved TRIGger:BRATe plus *RCL remote command causing application hang
- Fixed MTES:ALIG causing release assertion
- Resolved "DIG FUNCTioN2 and :SYSTem:MODE: JITT" being sent causing an application hang
- Resolved MEAS:CGR:JITT? command causing application exit
- Addressed :DISP:SCOL? CGR1 causing application exit
- Fixed TDR:DCALib RPCalib sequence creating application error
- Repaired ACQuire:RUNTil? CHANnel<N> query causing time out
- Repaired VIEW:RESP4 Crashing in Eye Mode
- Fixed MEASure:DELTatime? CHANnel3,FUNCTioN1 causing a crash
- Resolved :MEASure:APOWer?' command causing application crash

Software Option Packages (#200 and #201) Defects Fixed

- Addressed Total Jitter and Deterministic Jitter being optimistic by 1/10000 UI. The TJ histogram data had been missing the rightmost bin. This results in an optimistic estimate of TJ and DJ by up to 0.0001 UI, or about 10fs at 10GB/s (#200).
- Repaired TJ measurement being incorrectly identified in comma separated file. Fixed RJ value is also now indicated in exported jitter CSV file (#200).
- Resolved Periodic Jitter (PJ) Shape plot displaying bad data when data rate is an odd number. Previously, the result was a noisy trace (#200).
- Disabled Frequency resolve feature from working when jitter measurement results are questionable (#200).
- Repaired MatLab scripts without pipeline causing horizontal quantization in data display (#201).
- Enabled average power measurements when either equalizer or MatLab feature is enabled (#201).
- Fixed incorrect time base scale wrong with precision time base and equalizer in Eye mode. The time base for a channel is now scaled correctly when precision time base and equalizer are activated (#201).
- Resolved MatLab script and switch to TDR mode causing software crash. Added a check to verify whether or not the waveform being measured has any points at all (#201).

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.04.11 Release Notes, December 10, 2004 (Differences from Release A.04.10)

New Feature List

- None

Defects Fixed

- Repaired Autoscale function resulting in 3 eyes being displayed on Graticule with 86105C reference receivers.
- Resolved TDR cannot set scale to Ohms after default setup on channels 2 and 4. Previously, only channels 1 and 3 could be set to Ohms.
- Fixed average power monitor offset showing up on some reference receiver monitor.

Added additional settling time when setting the gain of the average power monitor.

- Repaired Extinction ratio calibration creating offset when 86107A precision timebase is active. The dark level measured during extinction ratio cal is too high when the cal is performed with precision timebase on.
- Fixed O/E calibration failing on 86105C Option 100 modules. Added an additional signal path to allow for completion of calibration for 86105C #100.
- Added support for 86105C unfiltered path.

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.04.10 Release Notes, November 1, 2004 (Differences from Release A.04.01)

New Feature List

- Support for 86105C Reference Receiver Module
- Provide the Enhanced Jitter Analysis Software Package (#200), which includes all previous jitter separation measurement capability, demonstrated in #100.
- The Bathtub jitter feature consists of two parts: the ability to display a plot of the extrapolated bathtub curve against the measured TJ histogram and the ability to select the BER at which TJ is measured.
- Increase pattern length supported by jitter mode to 2^{16}
- The Jitter Frequency Analysis feature will provide insight into the frequency content of periodic jitter. It will analyze both jitter that is correlated to the data rate (sub-rate jitter) as well as jitter that is uncorrelated from the data rate (asynchronous PJ). The items of interest are: jitter frequency, jitter magnitude, and relationship of jitter frequency to bit rate (i.e. is it a sub-rate)
- Added a Frequency component shape graph. After frequency analysis is turned on, the user will be able to display the shape of the jitter associated with a particular frequency. This aids in the identification of complicated jitter waveforms such as square waves, which has a large number of harmonically related frequencies.
- Added Advanced Waveform Analysis (#201)
- Built-in Linear Feedforward Equalizer application with up to 15 digital tap filter values
- Integrated MATLAB interface to allow signal processing using MATLAB (customer must purchase and install their own MATLAB licenses separately)
- Allow the customer to save and recall pattern waveforms which are long single-valued waveforms that span multiple bits, up to the entire pattern length of 2^{23} . Up to 4096
- points can be saved for each bit, which allows the file to be exported to other software applications like Excel or MATLAB for further analysis
- Support for 86100C Mainframe

The following features are available only with 86100C mainframe:

- Added Windows XP Operating System Service Pack 2. We recommend that all existing 86100C with Service Pack 1 installed update to Service Pack 2 because the additional security capabilities.
- The following feature is also available for 86100C mainframe and Option 001 Wide Bandwidth Trigger. Pattern Lock ("Eyeline?h) trigger will work when an 86107A precision time base module installed.
- The following feature is available with 86100C mainframe with Option 001 and Option 100 Jitter Software. Enabled 86107A precision time base to work in jitter mode, which reduces

the noise floor of jitter measurements down to 200 fs in most circumstances.

- Enhancements for 86100A/B/C Mainframes: Allow the 86107A precision time base modules to operate over continuous frequency bands. No hardware upgrade needed to existing modules, but customers must either provide their own filter or a clean sine wave with low total harmonic distortion.

Defects Fixed

- Repaired remote rise time measurement always returns data from lowest channel
- Fixed Remote Command to measure Extinction Ratio Factor (:MEAS...:ERF)
- Resolved Differential TDR Inverted channel marker values reporting incorrect value
- Addressed Auto Scale incorrectly aligns in time when multiple waveforms present on screen. There is now an optional bit rate parameter for the ":AUToscale" remote command. If bit rate information provided remotely, autoscale is much more robust (and much quicker!). Note that this parameter only applies to NRZ eye diagrams and clock signals.
- Repaired start of mask test aligns immediately even without sufficient data. Alignment will now be postponed up to 5 seconds if there is not enough data to align.
- Added remote command for reading TDR excess reactance. Added :MARKer:REACtance query command. It returns the value as follows: <reactance_value>,<units> where reactance value is in scientific notation and units are F or H. When there is no reactance value, a value of zero is returned and units of F.
- Repaired the remote command Disk:SimAge to Network drive breaking after excessive calls because a network drive resource was being opened and not closed.
- Fixed error in Scope Measurement - :MEASure:DELTAtime [source][,source]]. Instrument behavior is now consistent with documentation.
- Resolved Channel Input-Autoscale failed to send error message on marginally large signals. The correct notification occurs now for clipped signals.
- Repaired Horizontal scale being incorrect for clock signals while in jitter mode by adjusting the scales properly.
- Repaired remote command :System:DSP cmd with parameter '%' or '?' does't seem to work by fixing the instrument parser to recognize some more special characters.
- Fixed color pallet being incorrect when screens are saved in .JPG mode.
- Resolved ?eControl at limit error?f showing up when loading old setups by fixing the setting with the screen savers. There was a difference between 86100A/B and 86100C setups.
- Resolved Scope Measurement - :MEASure:TMIN returns Err: -224,Illegal parameter value.
- Repaired Channel Input - :CHAN:FDES? does not return Headers when :system:header on
- Changed default gain settings for 86117A Module Initialization to improve field servicing
- Fixed remote command Trig:Plock:Autodetect turns off pattern lock if autodetect fails
- Corrected Color grade incorrect when loading database memory in oscilloscope mode
- Repaired Histogram - Measure:Histogram:ppos?returns 9.99999E+37 while in TDR mode
- Fixed STM64_OC192SuperFEC_12_5.msk mask file. It should be a scaled version of the STM-64 OC-192 mask (rectangular, not hexagonal).
- *OPC now operates properly in Jitter Mode. *OPC, OPC? and *WAI should work properly now when switching to jitter mode using :SYStem:MODE JITTer or :DISK:LOAD "filename.jd",JDMemory.
- Disable mainframe ID check so 86110A is never reported when the instrument is queried
- Precision time bases 86107 using in geographies with Eastern time zones were reporting incorrect timebases. Setting the precision timebase time reference is limited in its repetition rate when pressing the dialog button. The way this timing is done has been changed so that it is no longer affected by changes to the system clock. Improved autoscaling with small signal levels
- 83495A Clock Recovery user interface for low loop BW should say 300 kHz not 30 kHz
- Repaired differential TDT normalized individual response looking like wrong scale

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.04.01 Release Notes, June 23, 2004 (Differences from Release A.04.00)

New Feature List

- None

Defects Fixed

- Fixed TDR measurement issue that was present on the 86100C only. Added 250 uS delay to allow step generator to settle.
- Fixed a modal dialog box defect on 86100C. If there was no trigger signal applied while in jitter mode, autoscale would proceed. Instead of reporting an autoscale failure due to an invalid trigger signal, it was possible that the "pattern lock lost" dialog would appear in an endless loop.
- Fixed an 86100C jitter mode sequence that could cause the software to lock up.

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.04.00 Release Notes, February 19, 2004 (Differences from Release A.03.05)

New Feature List

- Extinction Ratio correction factor
- Support for 83495A Clock Recovery Module
- Added software licensing. Permits the installation of optional software features.
- Support for 86100C Mainframe

The following features are available only with 86100C mainframe

- Added Windows XP Operating System. This allows user to access Windows features and applications.
- Disk recovery from D: partition. Allows recovery from software crash of main C: partition without external CD
- Uses standard Windows Installer to upgrade software

The following features are available only with 86100C mainframe and Option 001 Wide Bandwidth Trigger

- Pattern Lock ("Eyeline" trigger). Will display single valued waveforms for repetitive patterns.
- Auto-detection of pattern parameters. Relieves user of task of entering length and bit rate manually.

The following features are available with 86100C mainframe with Option 001 and Option 100 Jitter Software

- Added Jitter Mode. This is a fourth mode for the instrument that provides measurements

- of timing jitter.
- Separation and display of jitter components. Adds the pull down graphical display area.
- Save and recall of jitter memory. For use with jitter analysis data.

Defects Fixed

- Improved Touch Screen calibration for 86100C
- Fixed a debug assertion related to 54754A differential reference plane calibration
- Fixed a problem with saving an Instrument setup file on top of an existing setup file
- Fixed TDR remote command problem: If the right module is set to ON4 and TDTdest is Channel 3 and you issue the command to turn on both 3 and 4 an assertion error happens. The exact command was :TDR4:STIM ON3AND4;RESP3:TDTDest CHANNEL1
- Fixed problem with normalization of 54753A TDR module
- Fixed incorrect differential normalization (risetime of second channel). This was only broken in revision 3.06.
- Fixed a function that returns whether the TDR stimulus is differential/common mode or not. The incorrect value was returned if two TDR modules were installed and the right module was differential or common mode.
- Fixed a problem when using two TDR modules. Previously, when the command TDR2:STIM ON2 was followed by the command TDR2:STIM OFF, the left module was correctly turned off. However the right module was then inaccessible.
- Fixed problem where differential TDT cal data was not being saved and recalled correctly
- Fixed defects that were causing normalization cal data to be invalidated unnecessarily
- Fixed problem where TDT reference plane was lost when switching between stimulus channels
- Fixed defect where instrument could hang with the following command sequence in TDR mode: *rst;*cls;;autoscale;*opc?
- Fixed a problem with repetitive remote loading of instrument setup files
- Fixed a defect in the code that loads TDR/TDT cal data files. TDT data was being loaded but not fully resolved as valid.
- The :DIGitize command now works correctly with RESPonse<N> parameters. To digitize differential responses the type of each response must be specified by turning on the response before executing the :DIGitize command.

Known Issues

- 86119A (Optical Sampling Oscilloscope) is designed to work on software versions A.03.06 and A.03.05 only.

A.03.06 Release Notes, October 30, 2003 (Differences from Release A.03.05)

New Feature List

- None

Defects Fixed

- Differential normalization was not working correctly because each stimulus channel was using a different risetime value. Both channels now share the risetime setting of the first channel.
- Stopped circular dependencies that were causing autoscale to hang.
- Function that returns if TDR stimulus is differential/common mode was broken when two TDR modules were installed and the right module had differential or common mode stimulus. This prevented the :DIGitize remote command from working.
- Set of valid choices for the other module is now set correctly when turning off stimulus of

- 2 only or 4 only. Differential TDT data was not being saved and recalled properly.
- Normalization cal data was being invalidated inconsistently and unnecessarily when TDR settings were changed.
- Differential normalization behavior was based on the current setting of the differential mode control and not on the setting at the time the cal was performed.
- Fixed a bug where you would get "Control is set to limit" error messages the first time you power up an instrument.
- Single-ended TDT normalization data is now recalled properly when loading a TDR/TDT cal file. Previously the normalization cal data was read from the file but not fully resolved as valid.
- Enhanced TDT normalization so that thru measurements from calibration are stored independently from TDR short measurements. This allows greater flexibility in switching between stimulus channels.
- Fixed threading synchronization problem that caused the instrument to hang under certain conditions when a setup was recalled remotely.
- The :DIGitize remote command now works properly with RESPonse<N> parameters. If the response is differential or common mode it must be turned on before executing the :DIGitize command.

A.03.05 Release Notes, April 21, 2003 (Differences from Release A.03.04)

New Feature List

- Added support for the 86119A Optical Sampling Oscilloscope
- Added support for the new N1022A probe adaptor. This probe adaptor provides support for the InfiniiMax 3.5 to 7 GHz probes. This includes the following Agilent probes: 1131A, 1132A, 1134A, 1156A, 1157A and 1158A.
- Added support for horizontal mask alignment recommendation as per ITU-T G.691 for STM16 mask testing. This allows the mask center region to be positioned optimally along the horizontal axis to achieve the minimum number of mask violations. Along with this implementation, the following mask file was added:
- STM016_G.691_V2.0.msk
- Added the following NRZ masks to support testing to the Fibre Channel Physical Interface (FC-PI) Rev 13 standard for the following data rates: 1063, 2125 and 4250 MB/s.
- The mask file names are as follows: FC1063_PI_R13_Dec01.msk, FC2125_PI_R13_Dec01.msk, FC4250_PI_R13_Dec01.msk

Defects Fixed

- Improved color contrast between mask regions when saving screens with Invert Background Color checked. Previously, the mask margins appeared to be the same color as the mask.
- Corrected problem when loading waveform memories, scale and offset limits were incorrect. They were too limited when channel attenuation was a large value.
- Corrected problem with dropouts on channels when in differential TDR mode.
- Corrected problem with dropouts using high bandwidth modules when in TDT mode.
- Corrected problem with signal response being clipped when performing a differential TDR normalization
- Corrected problem with multiple presses of Stop/Single button causing data acquisition to halt.
- Corrected problem with Autoscale not completing when Histograms are enabled and acquisition is stopped.
- Corrected documentation with remote Programmer?fs Guide. Had incorrect syntax for probe calibration. Indicated command was :CAL:PROB:CHAN<n>, should be :CAL:PROB

CHAN<N>.

- Corrected debug assertion when the remote query :TRIGger:GATed? was sent.
- Corrected problem with :HARDcopy:DPRinter 0 printing remote command halting application.

A.03.04 Release Notes, October 10, 2002 (Differences from Release A.03.03)

New Feature List

- Add new vertical mask alignment method as specified by the OFSTP-4A standard. Alignment method can be selected from either GUI or remote access. GUI selection is done through Eye Boundary tab in the Configure Measurements dialog. Vertical alignment can be performed as originally implemented using the entire display (Vtop/Vbase) or as specified by OFSTP-4A, using the eye boundary 1/0 levels
- Modified NRZ eye measurement algorithms for rising and crossing point to better detect sharper edges
- 86117A/B optimized gain settings for better performance
- Add support for the 86109B Opt K10 . An enhanced version of the 86109B O/E receiver with 40 GHz optical bandwidth and 50 GHz electrical bandwidth
- Add support for the 86116B . High speed O/E receiver with 65 GHz optical bandwidth and warranted 80 GHz electrical bandwidth, typically 90 GHz

Defects Fixed:

- Mask files 10GbE_10_3125_May02.msk and 10GbE_9_953_May02.msk modified to align to eye boundary to be compliant with OFSTP-4A standard
- Corrected a problem with histogram mean always returning invalid data
- Corrected problems with TDR and TDT calibrations, markers and interactions with 54754A, 54753A, 83484A and other modules
- Corrected a problem with response waveform not yielding excess reactance
- *RST was not resetting source for math functions to channel1
- *RST was returning Error 241 Hardware missing
- Pressing Default Setup with no modules was displaying "Entry was not a valid selection" in the Dynamic Text Area
- Corrected problem with software upgrade process, cycling power too quickly could prevent upgrade from completing

A.03.03 Release Notes, July 18, 2002 (Differences from Release A.03.03)

New Feature List

- Add warning when attempting to save a file to a full floppy disk
- Add RZ/NRZ Eye mode selection GUI button
- Renamed horizontal and vertical calibration dialog box to avoid confusion with the precision timebase module
- TDR Differential Normalization feature added
- New masks: Infiniband, 10 GbE, Xaui
- Changes to electrical channel offset DAC for 86101A and 86103A modules
- The following new modules are supported: 86105B, 86107A Opt 010, 86117A/B, 86118A

Defects Fixed

- Modified the RZ algorithm to take care of unsymmetrical edges
- Corrected a problem with mask file names with a zero instead of an '0' i.e. STM64_OC192 written as STM64_OC1920
- Acquisition Limits misbehave with long patterns is now fixed

86100A-Only Firmware Release Notes

A.03.01 Release Notes, February 15, 2002 (Differences from Release A.02.20)

New Feature List

- Ability to acquire multiple channels at once in Eye Mode
- Ability to make measurement and mask analysis on multiple eyes at the same time
- Measurement results tab has a new look: clear meas button replaced by ?gx?h on meas results tab and tab up/down button replaced by ?g?P?h on meas results tab
- Parametric mask creation
- Ohms/Division setting added for TDR mode
- Ability to annotate and delete individual measurements
- Add support for the following new modules: 86111U, 86111A, 86115B, 86113A, 86102A, 86103B, 86116A, 86107A, 83494A Options 103 and 107

Defects Fixed:

- When filter is turned on, autoscale switches filter when completed for no reason
- Trigger setting changes after autoscale
- System crash when non-TDR module is autoscaled in TDR mode
- Stop Menu item behaves as Stop/Single
- Source parameter is not being accepted with 'MEASure' subsystem commands
- Sending DISP:PERS INF in eye mode sets it to CGR
- Screen capture when in color grade saves files in grayscale
- Saving .jpg on full diskette results in 0 record length file
- Remote uploaded waveform memories are not stored on disk
- Remote command returns mainframe calibrated when it is not; added :CAL:REC CHAn<N> command
- Quickmeasure attempts to measure waveform memories in eye mode
- Querying waveform values without waveform displayed hung the GUI
- Probe calibration hangs up box after remote command. We no longer initiate optical cal if chan is electrical.
- Need more digits in read-out of X Markers at low time/div setting
- MTEST:START fails after running MTEST:LOAD on bad mask name, but displays no warning
- MTEST:SAVE .msk saves as '.msk.pcm' ? Measurement Results Tab drops in height when

changing modes

- Measurement limits do not work when restored from saved setups
- Mask test align failure does not cause an error in the queue
- Mask margins not working with CG-GS waveforms
- Markers could be moved outside of the window
- Incorrect max string length reported with remote SYST:DSP command
- Inconsistent behavior when saving a file to a full floppy
- Horizontal Scale/Division setting didn't do anything
- Can't remove unused measurement from measurement results tab
- Binblock download of versus function does not work
- Autoscale resets transducer conversion factors
- Autoscale hung when acquisition stopped before using autoscale
- Autoscale failure on math waveform turns on other waveforms
- :VIEW CGM does not work after BLANK if a measurement command is sent first
- :DISplay:DATA? returns inconsistent color scheme
- :DISK:LOAD returns error 46 if in Oscilloscope mode
- TDR fails vertical calibration if Threshold is set to "Units"
- RZ eye width algorithm seems to place L open too close to mean; modified algorithm for unsymmetrical edges
- Module Delta Temperature reflects Mainframe temperature, not module
- Module calibration time comes from 2 different sources; problem during daylight savings time change
- Measurement Limit Test still displays measurement after it's been turned off
- Histogram tab stops displaying histogram data after math function
- Delta Y now only shows value in Measurement Results tab when the units are the same
- Databases dump when switching between RZ & NRZ measurements
- Cannot select CGM as source with :WAVEFORM:SOURCE+B34 command
- Cannot change channel offset for differential TDR channels
- :WAVEFORM:DATA? Returns incorrect values in ASCII mode
- :MEAS:CGRADe:JITTER? Does not return correct values; wasn't using source specified in :MEAS:SOURCE
- :CHAN#:UNIT:ATT, :CHAN#:UNIT:OFF and :CHAN#:PROBE now blocked when stimulus is on

A.02.20 Release Notes, August 2, 2001 (Differences from Release A.02.10)

New Feature List

- Japanese Language Help
- Support for 86106B
- Support for 86109B
- Support for 86102U

Defects Fixed:

- Fixed defect that prevents the mask margins from being enabled when a setup file containing margin information is loaded.
- Fixed defect that forced users to reload masks after performing an autoscale while running mask test.
- The acquisition speed while eye measurements are turned on, which had been previously degraded by a factor of four or five, is now fixed.
- GPIB disk store goes to setup directory instead of specified directory, this is now fixed.
- Resolved a problem with the RZ remote query for contrast ratio which causes the

instrument to crash while in eye mode.

- The remote command mask load (MTES:LOAD) does not work in "C:\User Files\Directory", this defect has been resolved.
- Modules insufficiently inserted produce an error message in the mainframe that tell the user to reinsert with more force.

A.02.10 Release Notes, May 1, 2001 (Differences from Release A.02.01)

New Feature List

- New Return to Zero measurements
- Added math function MIN/MAX
- Added JPEG and TIFF File Support
- Added scope measurements Vavg, Tmin, Tmax, Tedge, Delta Time
- Added Histogram Measurement Peak
- Ability to Show Histograms Window Boundary at all times when Histograms feature is turned on
- Added remote command to save screen image to hard or networked drive
- Added remote command to save file in any supported format

Defects Fixed:

- DCA crashes after default setup while in TDR mode
- Remote command :MARKER:Y1POSITION? now correctly returns answer in volts or ohms
- Resolved measurement resolution issue, introduced in version A.02.01. In Eye Mode on electrical channels the vertical scale resolution was fixed at 10 mV.
- Fixed text on Limit Setup button doesn't reset
- Filter rate values displayed correctly now
- Instrument Hangs when trying to move Y-Axis Markers
- Waveform colors no longer changed when invert waveform background color is selected
- Mask save screen command no longer overwrites previous data
- The remote command :MTESt:SCALe:SOURce query now returns SCPI standard character text, e.g. "CHAN1"
- Fixed a release assertion problem when mapping a Sun Unix drive
- Disk error message addressed because file overwriting itself
- Long Filenames entered from File\Save menu causes problem
- Waveform memory2 tab highlighted but information is about Waveform memory1
- Resolved a problem with mask load & reset after print causes illegal op & shutdown
- Fixed Display Data Query in INVERT format & printout showing non-invert
- Measurement Results Duplicate entry in Setup & Info
- Fixed remote command :DISPlay:LABel causes DCA to freeze in some remote sequences
- Can't format floppy drive when full of waveform files
- Signal goes to the bottom of screen at various vertical scale settings i.e. 6.1mV/div
- Fixed remote command :WAV:FORM WORD;SOURC CGR;DATA?

Measurement Enhancements

Improved built in Jitter measurement. The algorithm now uses a smaller window (3 pixels high) about the crossing point of the eye.

The EYE measurement is based on the histogram database.

In the case of jitter measurement, we need a "window" with vertical boundaeres of [y1,y2] and horizontal [x1,x2]. The jitter value is simply derived from the histogram within this window, e.g. the jitter RMS is the standard deviation of the histogram. The smaller the window, the smaller the jitter.

In firmware versions previous to A.02.10, 5% of the [Ymax,Ymin] is used, assuming that the "crossing" position is Ycross,

then

$$y1 = Y_{\text{cross}} - 5\%(Y_{\text{max}} - Y_{\text{min}})/2 \text{ and } y2 = Y_{\text{cross}} + 5\%(Y_{\text{max}} - Y_{\text{min}})/2$$

(Ymax is the max value in the database , Ymin the minimum)

In this version, A.02.10, it is changed to:

$$y1 = Y_{\text{cross}} - 1, y2 = Y_{\text{cross}} + 2$$

Since $5\%(Y_{\text{max}} - Y_{\text{min}}) \gg 3$ in most cases, therefore the current Y window is much smaller, when leads into smaller jitter measurement results.

A.02.01 Release Notes, January 30, 2001 (Differences from Release A.02.00)

New Feature List

- None

Defects Fixed

- Fixed a defect that can cause instrument lock-up if the markers are used below 40uW/div and off the 1/2/5 sequence
- Fixed a defect that caused linearity correction problems in 83486A modules. This resulted in potential vertical accuracy problems with these modules
- Fixed a defect that incorrectly identified the 83486A module and options. This also affected all modules that had an option number

Factory Only Enhancement

- Relaxed the test limits for timebase linearity on the horizontal calibration routine to help with production yield. This change has no impact on instrument performance

A.02.00 Release Notes, November 29, 2000 (Differences from Release A.01.22)

New Feature List

- Added Time Domain Reflectometry capabilities. The Agilent 54753A and 54754A TDR plug-in modules will work with the 86100A.
- Added the following Scope Measurements - Vbase, Vtop, Frequency, Pos Width, Neg Width, Duty Cycle, Vamp Eye, Bit Rate and Jitter RMS and Jitter Peak to Peak.
- Changed the finest resolution of the timebase from 10ps/div to 2ps/div.
- Changed the wavelength limit on User Optical Calibration from 1600 to 1625nm; however functionality beyond 1600 nm is not warranted or supported.
- Added the ability to create custom masks on a PC and input them into the DCA through the file manager.
- Provided the ability to remotely program the clock recovery modules.
- Added Zoom Box, the ability to zoom into a signal with multiple levels and to zoom back out.
- Added a wait cursor and new message give visual feedback that something is happening when a user inserts a module into the chassis.
- The Marker Select dialog box now supports OK/Cancel functionality.

Defects Fixed

- Fixed a defect that was causing the trace to jump to the top of the graticule when vertical offset was increased (the trace was moved down the screen).
- Release assertion on accessing algorithm for ACVrms. The MeasID for ACVrms and its corresponding HelpID were not in the measurement help search list.
- Fixed issues with waveform memory dialog box.
- Fixed a defect where 40 GBit/s would not highlight in the timebase dialog, when the user would choose it via the BitRateCtrl.
- Fixed a defect where failure occurred while bringing up saved waveform from file.
- Fixed a defect where saving screen image in .ps or .eps format causes a crash. A data buffer wasn't being allocated large enough which caused the stack and memory to be corrupted.
- The :DISK commands: CDIRectory and DIRectory now accept the long form of the token WAVEforms. WAVEforms was misspelled in the yuc code.
- "Orphan" radio controls - single radio controls which are not grouped with any other controls - now appear as text labels. These "display only" radio controls do not respond to any user action.
- 83485B and 86106A modules were defaulting to incorrect bandwidth for optical channels. Changed code to default optical channels to high bandwidth. The user can still set to low bandwidth via remote interface only.
- BackLight will not go off. Fixed this defect by ensuring that the screen saver timer only resets in the event of an actual change in mouse coordinates.
- WAV:DATA? returns extra carriage return. Carriage return only appears at the end of the output stream.
- The title of the cal failed dialog specifies the type of cal that failed instead of a generic 'Calibration failed' title.
- Exceeding # of display labels hangs the instrument. When the user tries to create more than 32 labels from the remote, a dialog box is displayed. The user needs to click on the continue button to continue. Since the front panel is locked by the remote, the user cannot continue. The dialog was replaced with a message in the display and a remote error.
- Default Setup was generating hardware missing errors when clock recovery modules were installed.
- Module channels of non clock modules were not being enabled when a clock module was installed upon default setup or module install.
- Measurement queries repeated in fast succession caused invalid result to be returned.
- Resolved MEAS:RES returning bad data.
- Fixed issue where the electrical bandwidth for 86106A and 86109A modules are incorrect. Changed Module script to reflect the actual bandwidth. Made changes so that the correct IFGain is used for building the acquisition look through table.
- Remote command to create directories only accepts drive A: or User Files, resolved.
- Remote disk commands don't permit the use of mapped drives, resolved.
- Autoscale resets power units from dBm to Watts when Quick measure is on, resolved.
- Autoscale improved to accommodate more waveforms.

Factory Only Enhancements

- Improved timebase calibration routine by making the polynomial calibration internal. This allows field timebase calibration, prior revisions can not be supported

A.01.22 Release Notes, September 6, 2000 (Differences from Release A.01.21)

NOTE

This firmware release is not supported.

New Feature List

- None

Defects Fixed

- The eye crossing algorithm is now more tolerant of heavily filtered eye waveforms. The Start-Mask-Test will align to larger variety of eye waveforms. Eye Measurements, which rely on an accurate eye crossing point, will perform better on these filtered waveforms.

A.01.21 Release Notes, August 9, 2000 (Differences from Release A.01.20)

NOTE

This firmware release is not supported.

New Feature List

- None

Defects Fixed

- None

Factory-only Enhancement

- New routine added to increase the efficiency of the factory horizontal calibration process. This improvement does not change the functionality of the product.

A.01.20 Release Notes, July 11, 2000 (Differences from Release A.01.13)

NOTE

This firmware release is not supported.

New Feature List

- External Scaling for vertical channel and trigger
- LAN capability for printing and file sharing

Defects Fixed

- With external keyboard connected, pressing ALT F4 allowed access to Windows which is bad. This "feature" is now unavailable.
- Vertical cal temperature window is changed from >1 deg C to >5 deg C.
- Changed variable persistence default time to 300 ms.
- If the touchscreen is disabled and a message, such as "Autoscale Failed" is displayed, the instrument will appear to be locked up because there is no way to dismiss the Message Box. The fix was to re-enable the touchscreen before displaying the Message box and the

- disabling it after it has been dismissed.
- If a vertical calibration fails and the module is removed before dismissing the Message Box an illegal operation error occurs. The Fix is to check for the presence of the module before reading from it.
 - Timebase resolution changed from 3 digits to 4, giving the same behavior as the 83480A.
 - Remotely invoking an extinction ratio cal, where a vertical cal is required, brings up a dialog box that cannot be cleared. The dialog box is replaced with a status message in the status bar.
 - Remotely sending the TEDGE command caused various problems. The fix is to only allow TEDGE? query and to reject TEDGE command.
 - Remotely selecting vertical markers from a time waveform to a versus waveform caused a "data out of range" error message; this has been fixed.
 - Invalid dates remotely sent to the instrument cause several problems, which are now resolved.

A.01.13 Release Notes, May 2, 2000 (Differences from Release A.01.12)

NOTE

This firmware release is not supported.

New Feature List

- None

Defects Fixed

- User Vertical Calibration routine that causes vertical calibration failures

